

XMC Input 100-240Vac, 600W Medical 7.5A Four Slot
XRA Input 100-240Vac, 200W Lo-noise Medical 4.5A Four Slot
XRB Input 100-240Vac, 400W Lo-noise Medical 5.5A Four Slot
XVA Input 100-240Vac, 400W Medical 7.5A Six Slot
XVB Input 100-240Vac, 700W Medical 9.5A Six Slot
XVC Input 100-240Vac, 1000W Medical 11.5A Six Slot
XVD Input 100-240Vac, 1200W Medical 11.5A Six Slot
XZA Input 100-240Vac, 400W Lo-noise Medical 7.5A Six Slot
XZB Input 100-240Vac, 900W Lo-noise Medical 11.5A Six Slot
See Enclosure 7-07, for detailed powerPac ratings.

Xgen powerMod Modules:

Xg1 Output 1.5 - 3.6Vdc 50A max 125/100W max
Xg2 Output 3.2 - 6.0Vdc 40A max 200/150W max
Xg3 Output 6.0 - 15.0Vdc 20A max 240/180W max
Xg4 Output 12.0 - 30.0Vdc 10A max 240/180W max
Xg5 Output 28.0 - 58.0Vdc 6A max 288/215W max
Xg6 Not used.
Xg7 Output 5.0 - 28.0Vdc 4A max 96/72W max
Xg8 Output (Dual) 5/5 - 28/28Vdc 2/2A max 48/48-36/36W max
See Enclosure 7-08, for detailed powerMod ratings.

Standards:

UL 60601-1, First Edition (2003)
CAN/CSA-C22.2 No.601.1-M90 with updates 1 and 2

Applicant Name and
Address:

EXCELSYS TECHNOLOGIES LTD
SWORDS BUSINESS PARK
DUBLIN IRELAND

This Report includes the following parts, in addition to this cover page:

1. Specific Inspection Criteria
2. Specific Technical Criteria
3. Clause Verdicts
4. Critical Components
5. Test Results
6. National Differences
7. Enclosures

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of Underwriters Laboratories Inc. ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

UL authorizes the applicant to reproduce the latest pages of the referenced Test Report consisting of the first page of the Specific Technical Criteria through to the end of the Conditions of Acceptability.


Any information and documentation provided to you involving UL Mark services are provided on behalf of Underwriters Laboratories Inc.

Test Report By:



Paul Daysh
Project Engineer
UL International (UK) Limited

Reviewed By:



Dan Bejnarowicz
Manager
UL International (UK) Limited